



**REQUEST FOR QUOTATION (RFQ) FOR PROCUREMENT OF PULSE OXIMETER FOR
PEADS DIGITAL**

RFQ Ref. # 01-300921R001-Sept-2021

Dear Concern,

Pakistan Kidney and Liver Institute & Research Center is a public hospital that specializes in Liver and Kidney transplants, serving the masses with state of the art facility while giving financial assistance to the deserving patients.

For following demand in our institute, please share quotation with best possible rates by **11-October- 2021** not later than **02:00pm**.

S. No	Description	QTY
1.	Pulse oximeter for Peads Small Description: 1.Lithium battery design, convenient charging 2.Smaller size, suitable for children testing 3. The blood Oxygen, The Pulse, The bar chart, The pulse wave form display 4. Detail description and Specification are attached.	04

Quotations are required only from sales tax registered firms.

1. PKLI is exempt from GST/PST.
2. The quotation should be printed on company letterhead and duly signed & stamped.
3. The quotation is evaluated on a composite rate basis.
4. Only one brand must be quoted otherwise quotation will be rejected.
5. Brand should be mentioned against each item.
6. Original sealed quotation must be delivered in the Mail Section of PKLI by **11-October -2021** not later than **02:00 pm. (Quotation will not be entertained through email)**
7. Completion is required within 04 days after issuance of Purchase Order.
8. In case of any/further clarification please contact.
9. Samples will be required on demand by OPD Department PKLI (If any).
10. For any technical inquiry/assistant please contact Mr. Ahsan Zafar (03087926636).
11. Quantity may be increased or decreased within the approved budget.

Looking forward to your prompt response.

Address:

HOD Procurement

Pakistan Kidney and Liver Hospital & Research Center
Package A, 1st Floor, One PKLI Avenue, Opposite DHA Phase-6, Knowledge City,
Lahore. Tel: 92-42-111-117-554, Mobile # 03087926636


X/ HSA
07/10/21

PRODUCT INFORMATION

DESCRIPTION:

Specification:

- SpO₂: Measurement range: 70%~99%
- Accuracy: $\pm 3\%$ on the stage of 70%~99%
- Resolution: $\pm 1\%$
- PR: Measurement range: 30BPM~240 BPM
- Accuracy: ± 2 BPM
- Automatically Power-off: The product automatically shuts off when no signal can be detected within 8 second



Jatadaw
2/10/2021
Dr. Jatadaw

Description:

- The Lovely Bird's exterior design, Children are easy to accept;
- Lithium battery design, convenient charging;
- Smaller size, suitable for children testing;
- Low power indicator;
- The Blood Oxygen, The Pulse, The bar chart, The pulse waveform Display;
- Automatic power off function, energy saving;
- No signal it will automatically power off in 8 seconds, energy saving.